



# Certificate of Conformity

**Certificate:** 13349-5S

**Master Contract:** LC 94447

**Project:** 13349

**Date Issued:** May 31, 2017

**Issued to:** **Critical Environment Technologies Canada Inc.**  
Unit 145, 7391 Vantage Way,  
Delta, British Columbia,  
V4G 1M3, Canada

**Certification System Type 3**  
(ISO / IEC Guide 67)

***The products listed below are eligible to bear the LabTest Certification Mark with adjacent indicators 'C' and 'US', and the subject area as shown below.***



**Issued by:**

  
Certification Manager

## **PRODUCTS**

- Gas Detectors
- Model Nos.: SCC-x-aaaa-aaaa-o & DCC-x-aaaa-bbbb-oo
- Ratings:90 – 240Vac, 50 – 60Hz / 16 – 30Vdc / 12 – 28Vac, 10W

## **APPLICABLE REQUIREMENTS & SUBJECT AREA**

- CSA-C22.2 No. 205-M1983 (R2009)
- UL508 (Edition 17): 2007

## **CONDITIONS OF ACCEPTABILITY**

- Permanently connected, wall mount, pollution degree II
- Intended to be used indoors and outdoors in ambient temperature of -40°C to 50°C.

The 'C' and 'US' indicators adjacent to the LabTest Certification Mark shall signify that the product has been evaluated by an accredited laboratory to the applicable CSA and UL Standards respectively. The subject area shall be displayed with the LabTest Certification Mark to identify the Standard Number, and Revision Level.